



日期  
Date: 2020.07.07

收件日期  
Date of Receipt: 2020.06.15

試驗報告  
TEST REPORT

土城場區  
TUCHENG

正本  
ORIGINAL

報告編號  
Report No.: TFF9F415

數量  
Quantity: 1件

報告頁次/頁數 (P2/3)  
Page Order/Pages: 來文字號 空白  
Ref. No.:

報告抬頭  
Report Title: 加利科技有限公司(T2460)

試件類別  
Item: 口罩

地址  
Address: 249 新北市八里區福德新村9號

試驗項目	試驗結果	試驗方法
細菌過濾效率(%)	99.4	CNS 14774-T5017-2018 9.2
金黃色葡萄球菌	99.5	CNS 14775 T4037-2003
ATCC 6538	99.4	
	99.3	
	99.3	
	99.3	

註：對照組的平均菌落數：2578 CFU。

註：平均粒徑：3.1  $\mu\text{m}$ 。

註：測試面：外側。

註：測試面積為39.5  $\text{cm}^2$ 。

註：測試流速：28.3 L/min。

註：依委託者所提供來樣資料為："加利" 醫用口罩 (未滅菌)

註：試驗報告僅就委託者之委託事項提供試驗結果，不對產品合法性做判斷。

附記: 1.本報告僅對樣品負責，樣品保留期限為一個月。

Note: This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.本報告非經本檢測及驗證部書面同意，不得摘錄複製，但全部複製除外。


This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.試驗報告所載事項不得作為公開廣告及商業推銷之用。

The test report should not be used for public advertisement and commercial promotion.

財團法人紡織產業綜合研究所  
所長授權核發人:

Authorized by president of  
Taiwan Textile Research Institute



紡織產業綜合研究所檢測及驗證部 新北市土城區承天路6號  
Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108, +886-2-22689839